

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL031124\
 Data File : VL041087.D
 Acq On : 11 Mar 2024 14:15
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/12/2024
 Supervised By :Mahesh Dadoda 03/12/2024

Quant Time: Mar 12 05:52:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL031124AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Tue Mar 12 05:47:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.198	49	1385346	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.664	114	3497943	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.481	117	3062886	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.796	95	2416662	10.092	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.900%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	2.915	62	3113m	0.040	ppbv	
32) 1,1,1-Trichloroethane	6.008	97	5578m	0.028	ppbv	
35) Carbon Tetrachloride	6.513	117	6125m	0.030	ppbv	
38) Trichloroethene	7.288	130	2431m	0.025	ppbv	
52) Tetrachloroethene	10.629	164	2014m	0.025	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.072	83	6387m	0.028	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL031124\
Data File : VL041087.D
Acq On : 11 Mar 2024 14:15
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Quant Time: Mar 12 05:52:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL031124AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Tue Mar 12 05:47:21 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 03/12/2024
Supervised By : Mahesh Dadoda 03/12/2024

