

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL031219\
Data File : VL033281.D
Acq On : 12 Mar 2019 12:31
Operator : SY/AP
Sample : VL0312ABL01
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0312ABL01

Quant Time: Mar 13 08:05:39 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL031219AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Mar 13 07:17:55 2019
QLast Update : Wed Mar 13 07:17:55 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	908470	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	2061313	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	2076666	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	1648310	10.28	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.80%

Target Compounds					Qvalue
15) Acetone	3.92	43	14701	0.105 ppbv	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL031219\
Data File : VL033281.D
Acq On : 12 Mar 2019 12:31
Operator : SY/AP
Sample : VL0312ABL01
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0312ABL01

Quant Time: Mar 13 08:05:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL031219AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Mar 13 07:17:55 2019
QLast Update : Wed Mar 13 07:17:55 2019
Response via : Initial Calibration

