

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL031219\
 Data File : VL033283.D
 Acq On : 12 Mar 2019 14:25
 Operator : SY/AP
 Sample : 0.4 PPB MDL
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 0.4 PPB MDL

Manual Integrations
 APPROVED

MMDadoda
 3/14/2019 5:06:00 AM

Quant Time: Mar 13 08:43:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL031219AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Mar 14 08:43:42 2019
 QLast Update : Wed Mar 13 07:17:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	869759	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	1958569	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	1924613	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	1516841	10.21	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
13) Ethanol	3.67	45	9209m	0.574	ppbv	
40) 1,4-Dioxane	8.00	88	11667m	0.456	ppbv	
66) Benzyl Chloride	17.15	91	26640m	0.275	ppbv	
80) Naphthalene,2-methyl-	25.12	142	4200	0.064	ppbv #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL031219\
Data File : VL033283.D
Acq On : 12 Mar 2019 14:25
Operator : SY/AP
Sample : 0.4 PPB MDL
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
0.4 PPB MDL

Manual Integrations
APPROVED

MMDadoda
3/14/2019 5:06:00 AM

Quant Time: Mar 13 08:43:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL031219AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Mar
QLast Update : Wed Mar 13 07:17:55 2019
Response via : Initial Calibration

