

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL031219\
 Data File : VL033284.D
 Acq On : 12 Mar 2019 15:03
 Operator : SY/AP
 Sample : 0.4 PPB MDL
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 0.4 PPB MDL

Manual Integrations
 APPROVED

MMDadoda
 3/14/2019 5:06:02 AM

Quant Time: Mar 13 08:58:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL031219AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Mar 14 08:58:36 2019
 QLast Update : Wed Mar 13 07:17:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	830198	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	1911116	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	1867982	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	1510737	10.47	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.70%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
13) Ethanol	3.67	45	10562m	0.690	ppbv	
40) 1,4-Dioxane	8.00	88	9486m	0.380	ppbv	
66) Benzyl Chloride	17.16	91	23672	0.252	ppbv	99
80) Naphthalene,2-methyl-	25.14	142	4157	0.066	ppbv #	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL031219\
Data File : VL033284.D
Acq On : 12 Mar 2019 15:03
Operator : SY/AP
Sample : 0.4 PPB MDL
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
0.4 PPB MDL

Manual Integrations
APPROVED

MMDadoda
3/14/2019 5:06:02 AM

Quant Time: Mar 13 08:58:36 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL031219AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Mar
QLast Update : Wed Mar 13 07:17:55 2019
Response via : Initial Calibration

