

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL031425\
 Data File : VL042148.D
 Acq On : 14 Mar 2025 16:29
 Operator : SY/MD
 Sample : QC031225 CAN 10300
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC031225 CAN 10300

Quant Time: Mar 14 23:19:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Feb 26 01:08:36 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	155212	10.000	ppbv	-0.01
33) 1,4-Difluorobenzene	3.962	114	353412	10.000	ppbv	-0.02
55) Chlorobenzene-d5	8.885	117	297273	10.000	ppbv	-0.02

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.380	95	230140	10.063	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.600%

Target Compounds						Qvalue
15) Acetone	1.839	43	7765	0.567	ppbv #	89
20) Methylene Chloride	2.062	84	2343	0.452	ppbv #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL031425\
Data File : VL042148.D
Acq On : 14 Mar 2025 16:29
Operator : SY/MD
Sample : QC031225 CAN 10300
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC031225 CAN 10300

Quant Time: Mar 14 23:19:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Feb 26 01:08:36 2025
Response via : Initial Calibration

