

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL031519\
 Data File : VL033330.D
 Acq On : 15 Mar 2019 13:00
 Operator : SY/AP
 Sample : VSTDIC0.1
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.1

Manual Integrations
 APPROVED

MMDadoda
 3/19/2019 10:02:23 AM

Quant Time: Mar 15 13:54:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL031519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Mar 15 2019
 QLast Update : Fri Mar 15 12:24:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	844814	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	1946924	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	2076156	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	1579684	10.01	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.30	62	5301m	0.099	ppbv	
32) 1,1,1-Trichloroethane	6.62	97	13457m	0.097	ppbv	
35) Carbon Tetrachloride	7.14	117	12206m	0.087	ppbv	
38) Trichloroethene	7.97	130	6874m	0.100	ppbv	
52) Tetrachloroethene	11.38	164	6950m	0.099	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.84	83	17966m	0.097	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL031519\
Data File : VL033330.D
Acq On : 15 Mar 2019 13:00
Operator : SY/AP
Sample : VSTDIC0.1
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.1

Manual Integrations
APPROVED

MMDadoda
3/19/2019 10:02:23 AM

Quant Time: Mar 15 13:54:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL031519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Mar
QLast Update : Fri Mar 15 12:24:28 2019
Response via : Initial Calibration

