

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL031519\
 Data File : VL033346.D
 Acq On : 16 Mar 2019 1:03
 Operator : SY/AP
 Sample : K1827-07DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 A-7DL

Manual Integrations
 APPROVED

MMDadoda
 3/19/2019 10:01:33 AM

Quant Time: Mar 16 06:38:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL031519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Mar 16 06:38:44 2019
 QLast Update : Fri Mar 15 14:52:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	871124	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	1944080	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	1835979	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	1403586	10.32	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.20%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.05	41	10517m	0.236	ppbv	
13) Ethanol	3.65	45	214085	10.061	ppbv	99
15) Acetone	3.91	43	94768	0.700	ppbv	98
19) Isopropyl Alcohol	4.03	45	82894m	1.287	ppbv	
25) Ethyl Acetate	5.77	43	41336	0.210	ppbv #	95
26) Hexane	5.78	57	13611	0.155	ppbv #	64
53) Toluene	9.95	91	244688	1.238	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL031519\
Data File : VL033346.D
Acq On : 16 Mar 2019 1:03
Operator : SY/AP
Sample : K1827-07DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
A-7DL

Manual Integrations
APPROVED

MMDadoda
3/19/2019 10:01:33 AM

Quant Time: Mar 16 06:38:44 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL031519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Mar
QLast Update : Fri Mar 15 14:52:14 2019
Response via : Initial Calibration

