

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL031620\
 Data File : VL034892.D
 Acq On : 16 Mar 2020 16:04
 Operator : SY/AP
 Sample : L1924-02DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA-02DL

Manual Integrations
 APPROVED

MMDadoda
 3/17/2020 12:15:53 PM

Quant Time: Mar 17 10:56:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL030920AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar 0
 QLast Update : Mon Mar 09 16:42:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	914135	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	1958907	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.10	117	1944503	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.44	95	1571199	10.36	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	2.98	41	10772m	0.221	ppbv	
13) Ethanol	3.57	45	57808	3.900	ppbv	93
15) Acetone	3.83	43	102062	0.861	ppbv	94
19) Isopropyl Alcohol	3.96	45	15397	0.206	ppbv #	49
26) Hexane	5.67	57	20816m	0.229	ppbv	
53) Toluene	9.81	91	51235	0.254	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL031620\
Data File : VL034892.D
Acq On : 16 Mar 2020 16:04
Operator : SY/AP
Sample : L1924-02DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA-02DL

Manual Integrations
APPROVED

MMDadoda
3/17/2020 12:15:53 PM

Quant Time: Mar 17 10:56:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL030920AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar
QLast Update : Mon Mar 09 16:42:53 2020
Response via : Initial Calibration

