

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL031620\
 Data File : VL034897.D
 Acq On : 16 Mar 2020 19:24
 Operator : SY/AP
 Sample : L1924-04DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OA-01DL

Manual Integrations
 APPROVED

MMDadoda
 3/17/2020 12:16:04 PM

Quant Time: Mar 17 11:04:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL030920AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar 0
 QLast Update : Mon Mar 09 16:42:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	870641	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	1863060	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.10	117	1873392	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.44	95	1490683	10.20	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.58	45	18484m	1.309	ppbv	
15) Acetone	3.84	43	44491m	0.394	ppbv	
53) Toluene	9.81	91	22879m	0.119	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL031620\
Data File : VL034897.D
Acq On : 16 Mar 2020 19:24
Operator : SY/AP
Sample : L1924-04DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OA-01DL

Manual Integrations
APPROVED

MMDadoda
3/17/2020 12:16:04 PM

Quant Time: Mar 17 11:04:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL030920AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar
QLast Update : Mon Mar 09 16:42:53 2020
Response via : Initial Calibration

