

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL031620\
 Data File : VL034905.D
 Acq On : 17 Mar 2020 00:38
 Operator : SY/AP
 Sample : L1924-06DL 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV-02DL

Quant Time: Mar 17 11:57:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL030920AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar 09 16:42:53 2020
 QLast Update : Mon Mar 09 16:42:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	854882	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	1806206	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.10	117	1762654	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.44	95	1427070	10.38	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.80%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
12) 1,1,2-Trichlorotrifluoroet	4.46	101	2474546	28.389	ppbv	99
15) Acetone	3.83	43	120535	1.088	ppbv	98
18) 1,1-Dichloroethene	4.26	96	385258	9.981	ppbv	98
24) 1,1-Dichloroethane	4.99	63	36153	0.329	ppbv	99
29) Chloroform	5.74	83	69752	0.538	ppbv	98
31) cis-1,2-Dichloroethene	5.53	61	104055	1.197	ppbv	96
32) 1,1,1-Trichloroethane	6.50	97	504358	3.991	ppbv	97
38) Trichloroethene	7.83	130	885393	14.480	ppbv	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL031620\
Data File : VL034905.D
Acq On : 17 Mar 2020 00:38
Operator : SY/AP
Sample : L1924-06DL 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV-02DL

Quant Time: Mar 17 11:57:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL030920AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar 09 16:42:53 2020
QLast Update : Mon Mar 09 16:42:53 2020
Response via : Initial Calibration

