


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL031620\
 Data File : VL034884.D
 Acq On : 16 Mar 2020 9:38
 Operator : SY/AP
 Sample : BFB
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL030920AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Mon Mar 09 16:42:53 2020
 Last Update : Mon Mar 09 16:42:53 2020



AutoFind: Scans 3621, 3622, 3623; Background Corrected with Scan 3597

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	25.1	134826	PASS
75	95	30	66	59.1	317610	PASS
95	95	100	100	100.0	537173	PASS
96	95	5	9	6.4	34309	PASS
173	174	0.00	2	0.3	1049	PASS
174	95	50	120	63.9	343296	PASS
175	174	4	9	7.8	26752	PASS
176	174	93	101	99.2	340714	PASS
177	176	5	9	6.3	21362	PASS