

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL031720\
 Data File : VL034918.D
 Acq On : 17 Mar 2020 16:44
 Operator : SY/AP
 Sample : L1952-01DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 AA-1DL

Quant Time: Mar 18 17:23:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL030920AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar 09 16:42:53 2020
 QLast Update : Mon Mar 09 16:42:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1029541	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	2212751	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.10	117	2209354	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.44	95	1828119	10.61	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	106.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.83	43	79778	0.598	ppbv	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL031720\
Data File : VL034918.D
Acq On : 17 Mar 2020 16:44
Operator : SY/AP
Sample : L1952-01DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
AA-1DL

Quant Time: Mar 18 17:23:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL030920AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar 09 16:42:53 2020
QLast Update : Mon Mar 09 16:42:53 2020
Response via : Initial Calibration

