

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL031720\
 Data File : VL034914.D
 Acq On : 17 Mar 2020 12:49
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Mar 18 07:12:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL030920AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar 09 16:42:53 2020
 QLast Update : Mon Mar 09 16:42:53 2020
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	106	0.00
2 T	Dichlorodifluoromethane	1.369	1.018	25.6	103	0.00
3	Chlorodifluoromethane	1.117	1.099	1.6	115	0.00
4	Chloromethane	0.631	0.637	-1.0	115	0.00
5 T	Vinyl Chloride	0.607	0.607	0.0	113	0.00
6 T	Bromomethane	0.353	0.349	1.1	113	0.00
7	Chloroethane	0.219	0.224	-2.3	118	0.00
8 T	Dichlorotetrafluoroethane	1.422	1.339	5.8	112	0.00
9 T	Propene	0.533	0.515	3.4	113	0.00
10 T	Heptane	1.313	1.339	-2.0	114	0.00
11 T	Trichlorofluoromethane	1.457	1.348	7.5	110	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.020	0.955	6.4	111	0.00
13	Ethanol	0.162	0.127	21.6	109	0.00
14 T	Bromoethene	0.440	0.428	2.7	110	0.00
15 T	Acetone	1.296	1.112	14.2	114	0.00
16 T	1,3-Butadiene	0.586	0.575	1.9	113	0.00
17	tert-Butyl alcohol	1.249	0.999	20.0	97	0.00
18 T	1,1-Dichloroethene	0.452	0.451	0.2	114	0.00
19 T	Isopropyl Alcohol	0.818	0.708	13.4	105	0.00
20 T	Methylene Chloride	0.410	0.392	4.4	114	0.00
21 T	Allyl Chloride	0.832	0.781	6.1	108	0.00
22 T	trans-1,2-Dichloroethene	0.458	0.442	3.5	111	0.00
23 T	Vinyl Acetate	1.915	1.854	3.2	110	0.00
24 T	1,1-Dichloroethane	1.283	1.241	3.3	112	0.00
25 T	Ethyl Acetate	2.309	2.233	3.3	113	0.00
26 T	Hexane	0.994	0.984	1.0	113	0.00
27 T	Carbon Disulfide	1.128	1.120	0.7	106	0.00
28 T	Methyl tert-Butyl Ether	1.654	1.518	8.2	104	0.00
29 T	Chloroform	1.517	1.475	2.8	114	0.00
30 T	Cyclohexane	0.830	0.824	0.7	111	0.00
31 T	cis-1,2-Dichloroethene	1.017	1.016	0.1	113	0.00
32 T	1,1,1-Trichloroethane	1.478	1.417	4.1	108	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
34 T	2-Butanone	0.665	0.673	-1.2	114	0.00
35 T	Carbon Tetrachloride	0.622	0.663	-6.6	108	0.00
36 T	Benzene	0.891	0.900	-1.0	111	0.00
37 T	1,2-Dichloroethane	0.503	0.525	-4.4	114	0.00
38 T	Trichloroethene	0.339	0.336	0.9	109	0.00
39 T	1,2-Dichloropropane	0.353	0.347	1.7	109	0.00
40 T	1,4-Dioxane	0.133	0.119	10.5	105	0.00
41 T	Tetrahydrofuran	0.360	0.386	-7.2	113	0.00
42 T	Bromodichloromethane	0.685	0.713	-4.1	109	0.00
43	Methyl Methacrylate	0.376	0.404	-7.4	113	0.00
44 T	2,2,4-Trimethylpentane	1.506	1.532	-1.7	112	0.00
45 T	t-1,3-Dichloropropene	0.441	0.458	-3.9	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL031720\
 Data File : VL034914.D
 Acq On : 17 Mar 2020 12:49
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Mar 18 07:12:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL030920AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar 09 16:42:53 2020
 QLast Update : Mon Mar 09 16:42:53 2020
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	cis-1,3-Dichloropropene	0.524	0.536	-2.3	105	0.00
47 T	1,1,2-Trichloroethane	0.362	0.368	-1.7	111	0.00
48 T	Dibromochloromethane	0.587	0.596	-1.5	104	0.00
49 T	Bromoform	0.523	0.527	-0.8	99	0.00
50 T	4-Methyl-2-Pentanone	0.957	0.997	-4.2	112	0.00
51 T	2-Hexanone	0.777	0.820	-5.5	111	0.00
52 T	Tetrachloroethene	0.353	0.319	9.6	104	0.00
53 T	Toluene	1.029	1.046	-1.7	110	0.00
54 T	1,2-Dibromoethane	0.546	0.559	-2.4	109	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	104	0.00
56	1,1,1,2-Tetrachloroethane	0.449	0.413	8.0	103	0.00
57 T	Chlorobenzene	0.782	0.715	8.6	108	0.00
58 T	Ethyl Benzene	1.354	1.283	5.2	108	0.00
59 T	m/p-Xylene	1.135	1.050	7.5	109	0.00
60 T	o-Xylene	1.108	1.035	6.6	110	0.00
61 T	Styrene	0.777	0.766	1.4	107	0.00
62	Isopropylbenzene	1.675	1.546	7.7	108	0.00
63 T	1,1,2,2-Tetrachloroethane	0.752	0.739	1.7	113	0.00
64	n-propylbenzene	0.436	0.406	6.9	106	0.00
65	tert-Butylbenzene	1.464	1.345	8.1	107	0.00
66 T	Benzyl Chloride	0.611	0.526	13.9	87	0.00
67	sec-Butylbenzene	2.030	1.917	5.6	108	0.00
68 S	1-Bromo-4-Fluorobenzene	0.780	0.770	1.3	105	0.00
69	p-Isopropyltoluene	1.702	1.611	5.3	107	0.00
70	n-Butylbenzene	1.737	1.709	1.6	111	0.00
71	2-Chlorotoluene	1.268	1.204	5.0	110	0.00
72 T	4-Ethyltoluene	1.308	1.255	4.1	108	0.00
73 T	1,3,5-Trimethylbenzene	1.225	1.178	3.8	109	0.00
74 T	1,2,4-Trimethylbenzene	1.262	1.214	3.8	109	0.00
75 T	1,3-Dichlorobenzene	0.761	0.699	8.1	107	0.00
76 T	1,4-Dichlorobenzene	0.759	0.703	7.4	107	0.00
77 T	1,2-Dichlorobenzene	0.737	0.686	6.9	106	0.00
78 T	Hexachloro-1,3-Butadiene	0.616	0.583	5.4	108	0.00
79 T	Naphthalene	0.984	1.183	-20.2	106	0.00
80 T	Naphthalene,2-methyl-	0.257	0.524	-103.9#	105	0.00
81 T	1,2,4-Trichlorobenzene	0.625	0.682	-9.1	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0