

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL032019\
 Data File : VL033394.D
 Acq On : 20 Mar 2019 17:49
 Operator : SY/AP
 Sample : 0.4 MDL
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 0.4 MDL

Manual Integrations
 APPROVED

MMDadoda
 3/25/2019 8:26:26 AM

Quant Time: Mar 24 16:10:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL031519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Mar 1
 QLast Update : Fri Mar 15 14:52:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	796363	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	1841439	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.24	117	1886871	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	1437859	10.28	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.80%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.66	45	14265	0.733	ppbv	96
40) 1,4-Dioxane	7.99	88	11197m	0.515	ppbv	
66) Benzyl Chloride	17.15	91	35068m	0.303	ppbv	
80) Naphthalene,2-methyl-	25.12	142	30868	0.258	ppbv	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL032019\
Data File : VL033394.D
Acq On : 20 Mar 2019 17:49
Operator : SY/AP
Sample : 0.4 MDL
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
0.4 MDL

Manual Integrations
APPROVED

MMDadoda
3/25/2019 8:26:26 AM

Quant Time: Mar 24 16:10:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL031519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Mar
QLast Update : Fri Mar 15 14:52:14 2019
Response via : Initial Calibration

