

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL032019\
 Data File : VL033397.D
 Acq On : 20 Mar 2019 19:47
 Operator : SY/AP
 Sample : K1975-01DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OA1DL

Manual Integrations
 APPROVED

MMDadoda
 3/25/2019 8:26:29 AM

Quant Time: Mar 21 13:16:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL031519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Mar 1
 QLast Update : Fri Mar 15 14:52:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	811189	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	1850653	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.24	117	1791515	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	1355907	10.21	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.10%

Target Compounds					Ovalue
13) Ethanol	3.66	45	20940	1.057 ppbv	96
15) Acetone	3.91	43	64398	0.511 ppbv	94
20) Methylene Chloride	4.41	84	35046	0.852 ppbv	92
26) Hexane	5.77	57	24609m	0.300 ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL032019\
Data File : VL033397.D
Acq On : 20 Mar 2019 19:47
Operator : SY/AP
Sample : K1975-01DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OA1DL

Manual Integrations
APPROVED

MMDadoda
3/25/2019 8:26:29 AM

Quant Time: Mar 21 13:16:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL031519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Mar
QLast Update : Fri Mar 15 14:52:14 2019
Response via : Initial Calibration

