

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL032019\
 Data File : VL033409.D
 Acq On : 21 Mar 2019 3:34
 Operator : SY/AP
 Sample : K1975-04DL 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV2DL

Manual Integrations
 APPROVED

MMDadoda
 3/25/2019 8:26:40 AM

Quant Time: Mar 22 06:41:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL031519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Mar 15 14:52:14 2019
 QLast Update : Fri Mar 15 14:52:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	1202459	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	2830903	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.24	117	1778443	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	1326815	10.07	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.70%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.04	41	14137m	0.229	ppbv	
15) Acetone	3.90	43	445154	2.382	ppbv	99
32) 1,1,1-Trichloroethane	6.61	97	427046	2.187	ppbv	99
34) 2-Butanone	5.35	43	39665	0.213	ppbv	95
52) Tetrachloroethene	11.37	164	1437009	13.615	ppbv	98
53) Toluene	9.95	91	61018	0.212	ppbv	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL032019\
Data File : VL033409.D
Acq On : 21 Mar 2019 3:34
Operator : SY/AP
Sample : K1975-04DL 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV2DL

Manual Integrations
APPROVED

MMDadoda
3/25/2019 8:26:40 AM

Quant Time: Mar 22 06:41:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL031519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Mar
QLast Update : Fri Mar 15 14:52:14 2019
Response via : Initial Calibration

