

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL032019\
 Data File : VL033414.D
 Acq On : 21 Mar 2019 8:58
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Quant Time: Mar 22 06:55:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL031519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Mar 15 14:52:14 2019
 QLast Update : Fri Mar 15 14:52:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	698795	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.28	114	1633891	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.24	117	2390830	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	1192705	6.73	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	67.30%

Target Compounds					Ovalue
52) Tetrachloroethene	11.37	164	1720076	28.236	ppbv 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL032019\
Data File : VL033414.D
Acq On : 21 Mar 2019 8:58
Operator : SY/AP
Sample : VIBLK
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Mar 22 06:55:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL031519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Mar 15 14:52:14 2019
QLast Update : Fri Mar 15 14:52:14 2019
Response via : Initial Calibration

