

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL032421\  
 Data File : VL036462.D  
 Acq On : 24 Mar 2021 10:51  
 Operator :  
 Sample : VL0324ABL01  
 Misc : 400ml/MSVOA L  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_L  
 ClientSampleId :

Quant Time: Mar 25 08:20:28 2021  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL030321AIR.M  
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA\_LWed Mar 03 14:18:24 2021  
 QLast Update : Wed Mar 03 14:18:24 2021  
 Response via : Initial Calibration

| Internal Standards      | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|-------------------------|-------|------|----------|-------|-------|----------|
| 1) Bromochloromethane   | 5.65  | 49   | 1671071  | 10.00 | ppbv  | -0.02    |
| 33) 1,4-Difluorobenzene | 7.16  | 114  | 3686938  | 10.00 | ppbv  | -0.02    |
| 55) Chlorobenzene-d5    | 12.07 | 117  | 3392174  | 10.00 | ppbv  | -0.02    |

System Monitoring Compounds

|                             |        |       |          |          |      |        |
|-----------------------------|--------|-------|----------|----------|------|--------|
| 68) 1-Bromo-4-Fluorobenzene | 14.41  | 95    | 2711423  | 9.72     | ppbv | -0.02  |
| Spiked Amount               | 10.000 | Range | 65 - 135 | Recovery | =    | 97.20% |

| Target Compounds       |      |    |        |       | Ovalue |
|------------------------|------|----|--------|-------|--------|
| 15) Acetone            | 3.86 | 43 | 54269m | 0.284 | ppbv   |
| 20) Methylene Chloride | 4.33 | 84 | 11213m | 0.174 | ppbv   |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL032421\  
Data File : VL036462.D  
Acq On : 24 Mar 2021 10:51  
Operator :  
Sample : VL0324ABL01  
Misc : 400ml/MSVOA L  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_L  
ClientSampleId :

Quant Time: Mar 25 08:20:28 2021  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL030321AIR.M  
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA\_L Wed Mar 03 14:18:24 2021  
QLast Update : Wed Mar 03 14:18:24 2021  
Response via : Initial Calibration

