

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL032422\
 Data File : VL038753.D
 Acq On : 24 Mar 2022 18:20
 Operator : SY/AP
 Sample : N2048-01DL2 20X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 C0P93DL2

Quant Time: Mar 26 09:55:32 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL032122AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar 21 17:09:26 2022
 QLast Update : Mon Mar 21 17:09:26 2022
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	808229	10.00	ppbv	0.02
33) 1,4-Difluorobenzene	7.18	114	2190225	10.00	ppbv	0.02
55) Chlorobenzene-d5	12.09	117	1902788	10.00	ppbv	0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	1443549	9.90	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.00%

Target Compounds					Qvalue
52) Tetrachloroethene	11.24	164	787091	12.72	ppbv 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL032422\
Data File : VL038753.D
Acq On : 24 Mar 2022 18:20
Operator : SY/AP
Sample : N2048-01DL2 20X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
C0P93DL2

Quant Time: Mar 26 09:55:32 2022
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL032122AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar 21 17:09:26 2022
QLast Update : Mon Mar 21 17:09:26 2022
Response via : Initial Calibration

