

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL032519\
 Data File : VL033445.D
 Acq On : 25 Mar 2019 20:50
 Operator : SY/AP
 Sample : K1560-14 0.06 PPB
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 LOQ-AIR-02-QT1-2019

Manual Integrations
 APPROVED

MMDadoda
 4/1/2019 1:22:55 AM

Quant Time: Mar 26 09:03:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar 26 09:03:23 2019
 QLast Update : Mon Mar 25 16:04:49 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	750719	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	1748225	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	1697434	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	1351963	10.16	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.29	62	2921m	0.065	ppbv	
32) 1,1,1-Trichloroethane	6.62	97	6366m	0.052	ppbv	
35) Carbon Tetrachloride	7.14	117	6467m	0.057	ppbv	
38) Trichloroethene	7.96	130	3254m	0.059	ppbv	
52) Tetrachloroethene	11.37	164	3299m	0.060	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.84	83	7852m	0.054	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL032519\
Data File : VL033445.D
Acq On : 25 Mar 2019 20:50
Operator : SY/AP
Sample : K1560-14 0.06 PPB
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
LOQ-AIR-02-QT1-2019

Manual Integrations
APPROVED

MMDadoda
4/1/2019 1:22:55 AM

Quant Time: Mar 26 09:03:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar
QLast Update : Mon Mar 25 16:04:49 2019
Response via : Initial Calibration

