

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL032519\
 Data File : VL033461.D
 Acq On : 26 Mar 2019 7:26
 Operator : SY/AP
 Sample : K1991-12DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OA-1DL

Manual Integrations
 APPROVED

MMDadoda
 4/1/2019 1:23:17 AM

Quant Time: Mar 27 02:20:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Mar 27 02:20:27 2019
 QLast Update : Tue Mar 26 16:00:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	695220	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	1561361	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	1565171	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	1209555	9.86	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.92	43	32384	0.387	ppbv	# 88
53) Toluene	9.94	91	22846m	0.150	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL032519\
Data File : VL033461.D
Acq On : 26 Mar 2019 7:26
Operator : SY/AP
Sample : K1991-12DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OA-1DL

Manual Integrations
APPROVED

MMDadoda
4/1/2019 1:23:17 AM

Quant Time: Mar 27 02:20:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar
QLast Update : Tue Mar 26 16:00:50 2019
Response via : Initial Calibration

