

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL032519\
 Data File : VL033441.D
 Acq On : 25 Mar 2019 17:22
 Operator : SY/AP
 Sample : VSTDICV010
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleID :
 ICVVL032519

Quant Time: Apr 01 03:34:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL032519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar 26 16:00:50 2019
 QLast Update : Tue Mar 26 16:00:50 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	106	0.00
2 T	Dichlorodifluoromethane	1.458	1.595	-9.4	138	0.00
3	Chlorodifluoromethane	1.151	1.000	13.1	99	0.00
4	Chloromethane	0.609	0.537	11.8	99	0.00
5 T	Vinyl Chloride	0.599	0.535	10.7	99	0.00
6 T	Bromomethane	0.385	0.337	12.5	99	0.00
7	Chloroethane	0.223	0.200	10.3	99	0.00
8 T	Dichlorotetrafluoroethane	1.554	1.361	12.4	101	0.00
9 T	Propene	0.483	0.434	10.1	99	0.00
10 T	Heptane	1.292	1.185	8.3	102	0.00
11 T	Trichlorofluoromethane	1.797	1.571	12.6	100	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.207	1.082	10.4	103	0.00
13	Ethanol	0.125	0.108	13.6	91	0.00
14 T	Bromoethene	0.482	0.435	9.8	98	0.00
15 T	Acetone	1.202	0.954	20.6	98	0.00
16 T	1,3-Butadiene	0.477	0.442	7.3	90	0.00
17	tert-Butyl alcohol	0.221	0.183	17.2	90	0.00
18 T	1,1-Dichloroethene	0.519	0.473	8.9	101	0.00
19 T	Isopropyl Alcohol	0.761	0.631	17.1	92	0.00
20 T	Methylene Chloride	0.510	0.419	17.8	105	0.00
21 T	Allyl Chloride	0.760	0.720	5.3	101	0.00
22 T	trans-1,2-Dichloroethene	0.541	0.505	6.7	105	0.00
23 T	Vinyl Acetate	1.711	1.669	2.5	102	0.00
24 T	1,1-Dichloroethane	1.300	1.167	10.2	101	0.00
25 T	Ethyl Acetate	2.231	2.027	9.1	102	0.00
26 T	Hexane	0.992	0.905	8.8	104	0.00
27 T	Carbon Disulfide	1.180	1.188	-0.7	102	0.00
28 T	Methyl tert-Butyl Ether	1.429	1.374	3.8	101	0.00
29 T	Chloroform	1.700	1.515	10.9	102	0.00
30 T	Cyclohexane	0.785	0.750	4.5	101	0.00
31 T	cis-1,2-Dichloroethene	0.986	0.942	4.5	101	0.00
32 T	1,1,1-Trichloroethane	1.619	1.486	8.2	99	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	105	0.00
34 T	2-Butanone	0.601	0.572	4.8	102	0.00
35 T	Carbon Tetrachloride	0.708	0.682	3.7	99	0.00
36 T	Benzene	0.839	0.809	3.6	103	0.00
37 T	1,2-Dichloroethane	0.562	0.503	10.5	99	0.00
38 T	Trichloroethene	0.316	0.323	-2.2	102	0.00
39 T	1,2-Dichloropropane	0.329	0.304	7.6	100	0.00
40 T	1,4-Dioxane	0.117	0.108	7.7	98	0.00
41 T	Tetrahydrofuran	0.330	0.327	0.9	102	0.00
42 T	Bromodichloromethane	0.766	0.729	4.8	100	0.00
43	Methyl Methacrylate	0.341	0.334	2.1	101	0.00
44 T	2,2,4-Trimethylpentane	1.466	1.368	6.7	102	0.00
45 T	t-1,3-Dichloropropene	0.418	0.456	-9.1	98	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL032519\
 Data File : VL033441.D
 Acq On : 25 Mar 2019 17:22
 Operator : SY/AP
 Sample : VSTDICV010
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL032519

Quant Time: Apr 01 03:34:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL032519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar 26 16:00:50 2019
 QLast Update : Tue Mar 26 16:00:50 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	cis-1,3-Dichloropropene	0.488	0.512	-4.9	100	0.00
47 T	1,1,2-Trichloroethane	0.373	0.343	8.0	101	0.00
48 T	Dibromochloromethane	0.630	0.622	1.3	99	0.00
49 T	Bromoform	0.601	0.562	6.5	73	0.00
50 T	4-Methyl-2-Pentanone	0.880	0.849	3.5	101	0.00
51 T	2-Hexanone	0.656	0.670	-2.1	101	0.00
52 T	Tetrachloroethene	0.317	0.314	0.9	99	0.00
53 T	Toluene	0.975	0.976	-0.1	100	0.00
54 T	1,2-Dibromoethane	0.584	0.556	4.8	100	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	105	0.00
56	1,1,1,2-Tetrachloroethane	0.458	0.407	11.1	100	0.00
57 T	Chlorobenzene	0.776	0.688	11.3	101	0.00
58 T	Ethyl Benzene	1.335	1.268	5.0	100	0.00
59 T	m/p-Xylene	1.193	1.052	11.8	88	0.00
60 T	o-Xylene	1.153	1.061	8.0	100	0.00
61 T	Styrene	0.792	0.784	1.0	99	0.00
62	Isopropylbenzene	1.633	1.452	11.1	96	0.00
63 T	1,1,2,2-Tetrachloroethane	0.849	0.742	12.6	101	0.00
64	n-propylbenzene	0.414	0.367	11.4	94	0.00
65	tert-Butylbenzene	1.439	1.311	8.9	99	0.00
66 T	Benzyl Chloride	0.669	0.722	-7.9	95	0.00
67	sec-Butylbenzene	2.047	1.840	10.1	97	0.00
68 S	1-Bromo-4-Fluorobenzene	0.784	0.742	5.4	101	0.00
69	p-Isopropyltoluene	1.667	1.534	8.0	97	0.00
70	n-Butylbenzene	1.766	1.623	8.1	97	0.00
71	2-Chlorotoluene	1.242	1.117	10.1	95	0.00
72 T	4-Ethyltoluene	1.297	1.204	7.2	97	0.00
73 T	1,3,5-Trimethylbenzene	1.241	1.152	7.2	98	0.00
74 T	1,2,4-Trimethylbenzene	1.330	1.202	9.6	98	0.00
75 T	1,3-Dichlorobenzene	0.822	0.709	13.7	98	0.00
76 T	1,4-Dichlorobenzene	0.801	0.704	12.1	96	0.00
77 T	1,2-Dichlorobenzene	0.780	0.693	11.2	97	0.00
78 T	Hexachloro-1,3-Butadiene	0.516	0.519	-0.6	132	0.00
79 T	Naphthalene	1.138	1.180	-3.7	96	0.00
80 T	Naphthalene,2-methyl-	0.382	0.506	-32.5#	93	0.00
81 T	1,2,4-Trichlorobenzene	0.631	0.600	4.9	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0