

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL032521\
 Data File : VL036474.D
 Acq On : 25 Mar 2021 4:33
 Operator : SY/AP
 Sample : VSTDIC0.1
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/26/2021
 Supervised By :Mahesh Dadoda 03/29/2021

Quant Time: Mar 25 09:26:26 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Thu Mar 25 07:52:30 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1678021m	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	3767499	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	3301368	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	2607878	9.61	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.26	62	13167m	0.144	ppbv	
32) 1,1,1-Trichloroethane	6.51	97	24215	0.094	ppbv	# 74
35) Carbon Tetrachloride	7.02	117	29100m	0.117	ppbv	
38) Trichloroethene	7.84	130	13342m	0.118	ppbv	
52) Tetrachloroethene	11.22	164	13369m	0.120	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.68	83	31465m	0.114	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL032521\
Data File : VL036474.D
Acq On : 25 Mar 2021 4:33
Operator : SY/AP
Sample : VSTDICC0.1
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Manual Integrations
APPROVED

Quant Time: Mar 25 09:26:26 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032521
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Thu Mar 25 07:52:30 2021
Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 03/26/2021
Supervised By :Mahesh Dadoda 03/29/2021

