

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL032521\
 Data File : VL036475.D
 Acq On : 25 Mar 2021 5:10
 Operator : SY/AP
 Sample : VSTDICC0.03
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/26/2021
 Supervised By :Mahesh Dadoda 03/29/2021

Quant Time: Mar 25 09:28:30 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Thu Mar 25 07:52:30 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1600836m	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	3543467	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	3151777	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	2516660	9.72	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.20%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.25	62	3781m	0.043	ppbv	
32) 1,1,1-Trichloroethane	6.50	97	11486m	0.047	ppbv	
35) Carbon Tetrachloride	7.02	117	9744m	0.042	ppbv	
38) Trichloroethene	7.82	130	4936m	0.047	ppbv	
52) Tetrachloroethene	11.23	164	4192m	0.040	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.68	83	10408m	0.039	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL032521\
Data File : VL036475.D
Acq On : 25 Mar 2021 5:10
Operator : SY/AP
Sample : VSTDICC0.03
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Manual Integrations
APPROVED

Quant Time: Mar 25 09:28:30 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032521
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Thu Mar 25 07:52:30 2021
Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 03/26/2021
Supervised By :Mahesh Dadoda 03/29/2021

