

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL032521\
 Data File : VL036484.D
 Acq On : 25 Mar 2021 13:37
 Operator : SY/AP
 Sample : M1801-05DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA2DL

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 03/28/2021
 Supervised By : Mahesh Dadoda 03/29/2021

Quant Time: Mar 28 23:12:31 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Thu Mar 25 09:42:40 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1505406	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	3437969	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.08	117	3018554	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.41	95	2402606	9.99	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.02	41	36365	0.406	ppbv	88
13) Ethanol	3.60	45	346020m	32.406	ppbv	
15) Acetone	3.85	43	103433	0.698	ppbv #	67
20) Methylene Chloride	4.34	84	215409	3.105	ppbv	97
53) Toluene	9.80	91	56887m	0.168	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL032521\
Data File : VL036484.D
Acq On : 25 Mar 2021 13:37
Operator : SY/AP
Sample : M1801-05DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA2DL

Manual Integrations
APPROVED

Quant Time: Mar 28 23:12:31 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032521
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Thu Mar 25 09:42:40 2021
Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 03/28/2021
Supervised By :Mahesh Dadoda 03/29/2021

