

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL032521\
 Data File : VL036493.D
 Acq On : 25 Mar 2021 19:22
 Operator : SY/AP
 Sample : M1748-33DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA L
 ClientSampleId :
 GOI-SG-1(2021)DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/26/2021
 Supervised By :Mahesh Dadoda 03/29/2021

Quant Time: Mar 28 23:29:20 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Thu Mar 25 09:42:40 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1537778	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	3420627	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	3012794	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	2416718	10.06	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.61	45	55341m	5.074	ppbv	
15) Acetone	3.85	43	259993	1.718	ppbv #	89
20) Methylene Chloride	4.34	84	184920	2.610	ppbv	92
26) Hexane	5.68	57	207251	1.241	ppbv #	47

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL032521\
Data File : VL036493.D
Acq On : 25 Mar 2021 19:22
Operator : SY/AP
Sample : M1748-33DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
GOI-SG-1(2021)DL

Manual Integrations
APPROVED

Quant Time: Mar 28 23:29:20 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032521
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Thu Mar 25 09:42:40 2021
Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 03/26/2021
Supervised By :Mahesh Dadoda 03/29/2021

