

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL032621\
 Data File : VL036510.D
 Acq On : 26 Mar 2021 12:15
 Operator : SY/AP
 Sample : M1748-38DL2 40X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 GOI-SG-6(2021)DL2

Manual Integrations
 APPROVED

MMDadoda
 3/29/2021 4:38:37 PM

Quant Time: Mar 27 23:04:33 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar 27 2021
 QLast Update : Thu Mar 25 09:42:40 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.69	49	1500685	10.00	ppbv	0.02
33) 1,4-Difluorobenzene	7.20	114	3426821	10.00	ppbv	0.03
55) Chlorobenzene-d5	12.12	117	3013741	10.00	ppbv	0.03
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.45	95	2353249	9.80	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.00%
Target Compounds						
13) Ethanol	3.64	45	52874m	4.967	ppbv	Ovalue
38) Trichloroethene	7.86	130	1393752	11.332	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL032621\
Data File : VL036510.D
Acq On : 26 Mar 2021 12:15
Operator : SY/AP
Sample : M1748-38DL2 40X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
GOI-SG-6(2021)DL2

Manual Integrations
APPROVED

MMDadoda
3/29/2021 4:38:37 PM

Quant Time: Mar 27 23:04:33 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar
QLast Update : Thu Mar 25 09:42:40 2021
Response via : Initial Calibration

