

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL032621\
 Data File : VL036512.D
 Acq On : 26 Mar 2021 13:33
 Operator : SY/AP
 Sample : M1760-03DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SS-7-TROPHY-CABINETDL

Manual Integrations
 APPROVED

MMDadoda
 3/29/2021 4:38:45 PM

Quant Time: Mar 28 20:24:10 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar 28 2021
 QLast Update : Thu Mar 25 09:42:40 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.69	49	1505845	10.00	ppbv	0.02
33) 1,4-Difluorobenzene	7.20	114	3423548	10.00	ppbv	0.03
55) Chlorobenzene-d5	12.12	117	3009476	10.00	ppbv	0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.45	95	2347830	9.79	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.62	45	386130	36.152	ppbv	96
15) Acetone	3.87	43	216962	1.464	ppbv #	85
19) Isopropyl Alcohol	3.99	45	556957	6.831	ppbv #	96
20) Methylene Chloride	4.36	84	98668m	1.422	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL032621\
Data File : VL036512.D
Acq On : 26 Mar 2021 13:33
Operator : SY/AP
Sample : M1760-03DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SS-7-TROPHY-CABINETDL

Manual Integrations
APPROVED

MMDadoda
3/29/2021 4:38:45 PM

Quant Time: Mar 28 20:24:10 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar
QLast Update : Thu Mar 25 09:42:40 2021
Response via : Initial Calibration

