

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL032621\
 Data File : VL036529.D
 Acq On : 27 Mar 2021 1:08
 Operator : SY/AP
 Sample : M1760-17DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SS-5-TEACHERS-WORK-RMDL

Manual Integrations
 APPROVED

MMDadoda
 3/29/2021 4:40:08 PM

Quant Time: Mar 28 21:35:39 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar 28 2021
 QLast Update : Thu Mar 25 09:42:40 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1312751	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	2961857	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	2608025	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	2045381	9.84	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.60	45	141782m	15.227	ppbv	
15) Acetone	3.85	43	132006	1.022	ppbv	# 73
19) Isopropyl Alcohol	3.99	45	37450m	0.527	ppbv	
20) Methylene Chloride	4.34	84	84698m	1.400	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL032621\
Data File : VL036529.D
Acq On : 27 Mar 2021 1:08
Operator : SY/AP
Sample : M1760-17DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SS-5-TEACHERS-WORK-RMDL

Manual Integrations
APPROVED

MMDadoda
3/29/2021 4:40:08 PM

Quant Time: Mar 28 21:35:39 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar
QLast Update : Thu Mar 25 09:42:40 2021
Response via : Initial Calibration

