

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL032719\
 Data File : VL033482.D
 Acq On : 28 Mar 2019 00:09
 Operator : SY/AP
 Sample : K2042-15DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IAS-5-TEACHERS-ROOMDL

Manual Integrations
 APPROVED

MMDadoda
 4/1/2019 1:26:00 AM

Quant Time: Mar 28 09:54:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar 2
 QLast Update : Tue Mar 26 16:00:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.75	49	679368	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.28	114	1561435	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.23	117	1605882	10.00	ppbv	-0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	1259023	10.00	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.64	45	641249	75.445	ppbv	65
15) Acetone	3.90	43	63820	0.781	ppbv	92
19) Isopropyl Alcohol	4.02	45	153092	2.963	ppbv #	97
20) Methylene Chloride	4.41	84	11109m	0.320	ppbv	
53) Toluene	9.93	91	19640m	0.129	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL032719\
Data File : VL033482.D
Acq On : 28 Mar 2019 00:09
Operator : SY/AP
Sample : K2042-15DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IAS-5-TEACHERS-ROOMDL

Manual Integrations
APPROVED

MMDadoda
4/1/2019 1:26:00 AM

Quant Time: Mar 28 09:54:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar
QLast Update : Tue Mar 26 16:00:50 2019
Response via : Initial Calibration

