

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL032719\
 Data File : VL033464.D
 Acq On : 27 Mar 2019 10:20
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Mar 28 07:28:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Mar 26 16:00:50 2019
 QLast Update : Tue Mar 26 16:00:50 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	111	0.00
2 T	Dichlorodifluoromethane	1.458	1.314	9.9	119	0.00
3	Chlorodifluoromethane	1.151	1.088	5.5	113	0.00
4	Chloromethane	0.609	0.580	4.8	111	0.00
5 T	Vinyl Chloride	0.599	0.573	4.3	111	0.00
6 T	Bromomethane	0.385	0.366	4.9	112	0.00
7	Chloroethane	0.223	0.208	6.7	108	0.00
8 T	Dichlorotetrafluoroethane	1.554	1.436	7.6	112	0.00
9 T	Propene	0.483	0.460	4.8	110	0.00
10 T	Heptane	1.292	1.237	4.3	112	0.00
11 T	Trichlorofluoromethane	1.797	1.642	8.6	110	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.207	1.135	6.0	113	0.00
13	Ethanol	0.125	0.134	-7.2	117	0.00
14 T	Bromoethene	0.482	0.470	2.5	111	0.00
15 T	Acetone	1.202	1.033	14.1	112	0.00
16 T	1,3-Butadiene	0.477	0.489	-2.5	104	0.00
17	tert-Butyl alcohol	0.221	0.227	-2.7	117	0.00
18 T	1,1-Dichloroethene	0.519	0.503	3.1	113	0.00
19 T	Isopropyl Alcohol	0.761	0.801	-5.3	123	0.00
20 T	Methylene Chloride	0.510	0.449	12.0	118	0.00
21 T	Allyl Chloride	0.760	0.744	2.1	109	0.00
22 T	trans-1,2-Dichloroethene	0.541	0.515	4.8	112	0.00
23 T	Vinyl Acetate	1.711	1.717	-0.4	110	0.00
24 T	1,1-Dichloroethane	1.300	1.213	6.7	110	0.00
25 T	Ethyl Acetate	2.231	2.110	5.4	111	0.00
26 T	Hexane	0.992	0.943	4.9	113	0.00
27 T	Carbon Disulfide	1.180	1.245	-5.5	111	0.00
28 T	Methyl tert-Butyl Ether	1.429	1.406	1.6	108	0.00
29 T	Chloroform	1.700	1.593	6.3	112	0.00
30 T	Cyclohexane	0.785	0.790	-0.6	112	0.00
31 T	cis-1,2-Dichloroethene	0.986	0.973	1.3	110	0.00
32 T	1,1,1-Trichloroethane	1.619	1.563	3.5	109	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	111	0.00
34 T	2-Butanone	0.601	0.582	3.2	109	0.00
35 T	Carbon Tetrachloride	0.708	0.730	-3.1	112	0.00
36 T	Benzene	0.839	0.837	0.2	112	0.00
37 T	1,2-Dichloroethane	0.562	0.534	5.0	111	0.00
38 T	Trichloroethene	0.316	0.335	-6.0	112	0.00
39 T	1,2-Dichloropropane	0.329	0.315	4.3	109	0.00
40 T	1,4-Dioxane	0.117	0.126	-7.7	120	0.00
41 T	Tetrahydrofuran	0.330	0.332	-0.6	109	0.00
42 T	Bromodichloromethane	0.766	0.750	2.1	109	0.00
43	Methyl Methacrylate	0.341	0.342	-0.3	109	0.00
44 T	2,2,4-Trimethylpentane	1.466	1.396	4.8	110	0.00
45 T	t-1,3-Dichloropropene	0.418	0.467	-11.7	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL032719\
 Data File : VL033464.D
 Acq On : 27 Mar 2019 10:20
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Mar 28 07:28:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Mar 26 16:00:50 2019
 QLast Update : Tue Mar 26 16:00:50 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	cis-1,3-Dichloropropene	0.488	0.530	-8.6	109	0.00
47 T	1,1,2-Trichloroethane	0.373	0.355	4.8	110	0.00
48 T	Dibromochloromethane	0.630	0.651	-3.3	109	0.00
49 T	Bromoform	0.601	0.597	0.7	82	0.00
50 T	4-Methyl-2-Pentanone	0.880	0.866	1.6	109	0.00
51 T	2-Hexanone	0.656	0.692	-5.5	110	0.00
52 T	Tetrachloroethene	0.317	0.328	-3.5	109	0.00
53 T	Toluene	0.975	1.001	-2.7	108	0.00
54 T	1,2-Dibromoethane	0.584	0.578	1.0	109	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	109	0.00
56	1,1,1,2-Tetrachloroethane	0.458	0.431	5.9	109	0.00
57 T	Chlorobenzene	0.776	0.731	5.8	111	0.00
58 T	Ethyl Benzene	1.335	1.346	-0.8	110	0.00
59 T	m/p-Xylene	1.193	1.144	4.1	99	-0.03
60 T	o-Xylene	1.153	1.124	2.5	109	0.00
61 T	Styrene	0.792	0.838	-5.8	110	0.00
62	Isopropylbenzene	1.633	1.575	3.6	108	0.00
63 T	1,1,2,2-Tetrachloroethane	0.849	0.769	9.4	108	0.00
64	n-propylbenzene	0.414	0.406	1.9	107	0.00
65	tert-Butylbenzene	1.439	1.426	0.9	111	0.00
66 T	Benzyl Chloride	0.669	0.774	-15.7	105	0.00
67	sec-Butylbenzene	2.047	2.023	1.2	110	0.00
68 S	1-Bromo-4-Fluorobenzene	0.784	0.744	5.1	105	0.00
69	p-Isopropyltoluene	1.667	1.687	-1.2	110	0.00
70	n-Butylbenzene	1.766	1.789	-1.3	110	0.00
71	2-Chlorotoluene	1.242	1.221	1.7	107	0.00
72 T	4-Ethyltoluene	1.297	1.313	-1.2	109	0.00
73 T	1,3,5-Trimethylbenzene	1.241	1.262	-1.7	111	0.00
74 T	1,2,4-Trimethylbenzene	1.330	1.320	0.8	112	0.00
75 T	1,3-Dichlorobenzene	0.822	0.778	5.4	112	0.00
76 T	1,4-Dichlorobenzene	0.801	0.779	2.7	110	0.00
77 T	1,2-Dichlorobenzene	0.780	0.754	3.3	109	0.00
78 T	Hexachloro-1,3-Butadiene	0.516	0.412	20.2	108	0.00
79 T	Naphthalene	1.138	1.291	-13.4	108	0.00
80 T	Naphthalene,2-methyl-	0.382	0.560	-46.6#	107	0.00
81 T	1,2,4-Trichlorobenzene	0.631	0.659	-4.4	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0