


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL032719\
 Data File : VL033463.D
 Acq On : 27 Mar 2019 8:57
 Operator : SY/AP
 Sample : BFB
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL032519AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue Mar 26 16:00:50 2019
 Last Update : Tue Mar 26 16:00:50 2019



AutoFind: Scans 3617, 3618, 3619; Background Corrected with Scan 3594

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	23.7	160661	PASS
75	95	30	66	56.4	381674	PASS
95	95	100	100	100.0	676522	PASS
96	95	5	9	6.8	45786	PASS
173	174	0.00	2	0.9	4275	PASS
174	95	50	120	67.3	455552	PASS
175	174	4	9	6.8	30893	PASS
176	174	93	101	94.7	431296	PASS
177	176	5	9	6.4	27530	PASS