

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL032919\
 Data File : VL033499.D
 Acq On : 29 Mar 2019 12:54
 Operator : SY/AP
 Sample : OC032719 CAN 10266
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC032719 CAN 10266

Manual Integrations
 APPROVED

MMDadoda
 4/1/2019 4:03:08 AM

Quant Time: Mar 30 01:53:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar 2
 QLast Update : Tue Mar 26 16:00:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	866292	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	2184614	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	2099473	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	1610242	9.78	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.80%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
4) Chloromethane	3.17	50	11962	0.227	ppbv #	80
5) Vinyl Chloride	3.29	62	6209	0.120	ppbv	91
8) Dichlorotetrafluoroethane	3.22	85	13389	0.099	ppbv	99
20) Methylene Chloride	4.40	84	56691	1.282	ppbv	88
26) Hexane	5.78	57	34529	0.402	ppbv #	50
29) Chloroform	5.84	83	15550	0.106	ppbv	95
32) 1,1,1-Trichloroethane	6.62	97	13814	0.098	ppbv	98
35) Carbon Tetrachloride	7.13	117	14332m	0.093	ppbv	
36) Benzene	7.00	78	17972	0.098	ppbv #	94
38) Trichloroethene	7.96	130	7597m	0.110	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL032919\
Data File : VL033499.D
Acq On : 29 Mar 2019 12:54
Operator : SY/AP
Sample : OC032719 CAN 10266
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC032719 CAN 10266

Manual Integrations
APPROVED

MMDadoda
4/1/2019 4:03:08 AM

Quant Time: Mar 30 01:53:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar
QLast Update : Tue Mar 26 16:00:50 2019
Response via : Initial Calibration

