

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL032919\
 Data File : VL033501.D
 Acq On : 29 Mar 2019 14:14
 Operator : SY/AP
 Sample : K2042-16DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IAS-4-STORAGE-RM-BY-THE-CAFE

Quant Time: Mar 30 01:58:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar 26 16:00:50 2019
 QLast Update : Tue Mar 26 16:00:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	809819	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	2145414	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	2113153	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	1644037	9.92	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.20%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.64	45	206667	20.398	ppbv	64
15) Acetone	3.90	43	82938	0.852	ppbv	96
19) Isopropyl Alcohol	4.04	45	56964	0.925	ppbv #	90
20) Methylene Chloride	4.41	84	80702	1.952	ppbv	97
26) Hexane	5.78	57	64296	0.800	ppbv #	41
53) Toluene	9.95	91	68980	0.330	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL032919\
Data File : VL033501.D
Acq On : 29 Mar 2019 14:14
Operator : SY/AP
Sample : K2042-16DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IAS-4-STORAGE-RM-BY-THE-CAFE

Quant Time: Mar 30 01:58:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar 26 16:00:50 2019
QLast Update : Tue Mar 26 16:00:50 2019
Response via : Initial Calibration

