

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL032919\
 Data File : VL033514.D
 Acq On : 29 Mar 2019 22:52
 Operator : SY/AP
 Sample : K2042-03DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SG-6-RM-104DL

Manual Integrations
 APPROVED

MMDadoda
 4/1/2019 4:03:21 AM

Quant Time: Mar 31 15:06:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Mar 2
 QLast Update : Tue Mar 26 16:00:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	839488	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	1891773	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	1821540	10.00	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.58	95	1403232	9.83	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.66	45	467874	44.547	ppbv	66
15) Acetone	3.90	43	162992	1.615	ppbv	100
20) Methylene Chloride	4.41	84	103036	2.405	ppbv	93
26) Hexane	5.78	57	110216	1.324	ppbv #	39
38) Trichloroethene	7.96	130	6309m	0.106	ppbv	
53) Toluene	9.94	91	18127	0.098	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL032919\
Data File : VL033514.D
Acq On : 29 Mar 2019 22:52
Operator : SY/AP
Sample : K2042-03DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SG-6-RM-104DL

Manual Integrations
APPROVED

MMDadoda
4/1/2019 4:03:21 AM

Quant Time: Mar 31 15:06:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar
QLast Update : Tue Mar 26 16:00:50 2019
Response via : Initial Calibration

