

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL033021\
 Data File : VL036573.D
 Acq On : 30 Mar 2021 17:20
 Operator : SY/AP
 Sample : M1797-03DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 MF-SG-04(2021)DL

Manual Integrations
 APPROVED

MMDadoda
 4/1/2021 10:58:35 AM

Quant Time: Mar 31 21:02:03 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar 25 2021
 QLast Update : Thu Mar 25 09:42:40 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1409730	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.16	114	3323286	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.07	117	2937887	10.00	ppbv	-0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.39	95	2342095	10.00	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.59	45	458046m	45.809	ppbv	
15) Acetone	3.84	43	222354	1.603	ppbv #	79
20) Methylene Chloride	4.33	84	259507	3.995	ppbv	97
26) Hexane	5.67	57	440410	2.877	ppbv #	47

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL033021\
Data File : VL036573.D
Acq On : 30 Mar 2021 17:20
Operator : SY/AP
Sample : M1797-03DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
MF-SG-04(2021)DL

Manual Integrations
APPROVED

MMDadoda
4/1/2021 10:58:35 AM

Quant Time: Mar 31 21:02:03 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar
QLast Update : Thu Mar 25 09:42:40 2021
Response via : Initial Calibration

