

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL033021\
 Data File : VL036587.D
 Acq On : 31 Mar 2021 2:52
 Operator : SY/AP
 Sample : M1797-10DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 MF-SG-13(2021)DUPDL

Manual Integrations
 APPROVED

MMDadoda
 4/1/2021 10:59:50 AM

Quant Time: Mar 31 21:51:33 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar 25 2021
 QLast Update : Thu Mar 25 09:42:40 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1259360	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.16	114	2984108	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.07	117	2670473	10.00	ppbv	-0.03

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.40	95	2041158	9.59	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	95.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.60	45	181323m	20.299	ppbv	
15) Acetone	3.84	43	207154	1.671	ppbv #	76
20) Methylene Chloride	4.33	84	219235	3.778	ppbv	98
26) Hexane	5.67	57	402679	2.945	ppbv #	47

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL033021\
Data File : VL036587.D
Acq On : 31 Mar 2021 2:52
Operator : SY/AP
Sample : M1797-10DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
MF-SG-13(2021)DUPDL

Manual Integrations
APPROVED

MMDadoda
4/1/2021 10:59:50 AM

Quant Time: Mar 31 21:51:33 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar
QLast Update : Thu Mar 25 09:42:40 2021
Response via : Initial Calibration

