

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL033121\
 Data File : VL036592.D
 Acq On : 31 Mar 2021 6:26
 Operator : SY/AP
 Sample : VL0331ABL01
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0331ABL01

Quant Time: Apr 03 05:31:07 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Thu Mar 25 09:42:40 2021
 QLast Update : Thu Mar 25 09:42:40 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1257882	10.00	ppbv	-0.01
33) 1,4-Difluorobenzene	7.16	114	2885939	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.07	117	2511808	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.40	95	1967338	9.83	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.30%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL033121\
Data File : VL036592.D
Acq On : 31 Mar 2021 6:26
Operator : SY/AP
Sample : VL0331ABL01
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0331ABL01

Quant Time: Apr 03 05:31:07 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar 25 09:42:40 2021
QLast Update : Thu Mar 25 09:42:40 2021
Response via : Initial Calibration

