

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL033121\
 Data File : VL036610.D
 Acq On : 31 Mar 2021 20:27
 Operator : SY/AP
 Sample : M1808-02DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 AASG2DL

Manual Integrations
 APPROVED

MMDadoda
 4/2/2021 6:52:59 PM

Quant Time: Apr 02 15:22:38 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar 25 2021
 QLast Update : Thu Mar 25 09:42:40 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1694030	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	3750397	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.10	117	3332102	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.43	95	2636611	9.93	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.62	45	101719m	8.466	ppbv	
15) Acetone	3.87	43	132943	0.797	ppbv #	84
20) Methylene Chloride	4.35	84	215402	2.759	ppbv	90
26) Hexane	5.69	57	199618	1.085	ppbv #	45

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL033121\
Data File : VL036610.D
Acq On : 31 Mar 2021 20:27
Operator : SY/AP
Sample : M1808-02DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
AASG2DL

Manual Integrations
APPROVED

MMDadoda
4/2/2021 6:52:59 PM

Quant Time: Apr 02 15:22:38 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar
QLast Update : Thu Mar 25 09:42:40 2021
Response via : Initial Calibration

