

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL033121\
 Data File : VL036616.D
 Acq On : 1 Apr 2021 00:41
 Operator : SY/AP
 Sample : M1807-03DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SG3DL

Manual Integrations
 APPROVED

MMDadoda
 4/2/2021 6:53:39 PM

Quant Time: Apr 02 15:42:53 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar 25 2021
 QLast Update : Thu Mar 25 09:42:40 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	1612911	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	3753280	10.00	ppbv	0.01
55) Chlorobenzene-d5	12.11	117	3252435	10.00	ppbv	0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.44	95	2592063	10.00	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.63	45	77865m	6.806	ppbv	
15) Acetone	3.88	43	180375	1.136	ppbv #	86
19) Isopropyl Alcohol	4.02	45	48992m	0.561	ppbv	
20) Methylene Chloride	4.36	84	75411	1.015	ppbv #	90
26) Hexane	5.70	57	77148	0.440	ppbv #	44

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL033121\
Data File : VL036616.D
Acq On : 1 Apr 2021 00:41
Operator : SY/AP
Sample : M1807-03DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SG3DL

Manual Integrations
APPROVED

MMDadoda
4/2/2021 6:53:39 PM

Quant Time: Apr 02 15:42:53 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar
QLast Update : Thu Mar 25 09:42:40 2021
Response via : Initial Calibration

