

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL040221\
 Data File : VL036654.D
 Acq On : 2 Apr 2021 21:00
 Operator : SY/AP
 Sample : M1841-07DL2 40X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 PBS-SS-2DL2

Manual Integrations
 APPROVED

apatel
 4/5/2021 1:14:08 PM

Quant Time: Apr 04 16:49:33 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar 25 2021
 QLast Update : Thu Mar 25 09:42:40 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1284880	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.16	114	3338753	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.06	117	2968619	10.00	ppbv	-0.03

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
68) 1-Bromo-4-Fluorobenzene	14.39	95	2322485	9.82	ppbv	-0.03	
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.20%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.61	45	38445m	4.218	ppbv	
15) Acetone	3.84	43	568314m	4.494	ppbv	
19) Isopropyl Alcohol	3.98	45	348975m	5.016	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL040221\
Data File : VL036654.D
Acq On : 2 Apr 2021 21:00
Operator : SY/AP
Sample : M1841-07DL2 40X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
PBS-SS-2DL2

Manual Integrations
APPROVED

apatel
4/5/2021 1:14:08 PM

Quant Time: Apr 04 16:49:33 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar
QLast Update : Thu Mar 25 09:42:40 2021
Response via : Initial Calibration

