

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL040221\
 Data File : VL036656.D
 Acq On : 2 Apr 2021 22:19
 Operator : SY/AP
 Sample : M1853-04DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SOFF-SG-04DL

Manual Integrations
 APPROVED

apatel
 4/5/2021 1:14:11 PM

Quant Time: Apr 04 16:53:13 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar 25 2021
 QLast Update : Thu Mar 25 09:42:40 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1272714	10.00	ppbv	-0.03
33) 1,4-Difluorobenzene	7.15	114	3423769	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.06	117	2950035	10.00	ppbv	-0.03

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.39	95	2328316	9.90	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Chlorodifluoromethane	2.98	51	30183	0.127	ppbv	98
13) Ethanol	3.60	45	185535m	20.553	ppbv	
15) Acetone	3.84	43	129767	1.036	ppbv #	77
20) Methylene Chloride	4.33	84	177734	3.030	ppbv	97
26) Hexane	5.66	57	156979	1.136	ppbv #	47

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL040221\
Data File : VL036656.D
Acq On : 2 Apr 2021 22:19
Operator : SY/AP
Sample : M1853-04DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SOFF-SG-04DL

Manual Integrations
APPROVED

apatel
4/5/2021 1:14:11 PM

Quant Time: Apr 04 16:53:13 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar
QLast Update : Thu Mar 25 09:42:40 2021
Response via : Initial Calibration

