

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL040221\
 Data File : VL036660.D
 Acq On : 3 Apr 2021 00:51
 Operator : SY/AP
 Sample : M1853-03DL 40X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SOFF-SG-03DL

Manual Integrations
 APPROVED

apatel
 4/5/2021 1:14:17 PM

Quant Time: Apr 03 07:22:45 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar 25 2021
 QLast Update : Thu Mar 25 09:42:40 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1289874	10.00	ppbv	-0.03
33) 1,4-Difluorobenzene	7.15	114	3301292	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.06	117	2974330	10.00	ppbv	-0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.39	95	2299603	9.70	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.60	45	144164m	15.757	ppbv	
15) Acetone	3.85	43	63865m	0.503	ppbv	
19) Isopropyl Alcohol	3.99	45	171342m	2.453	ppbv	
20) Methylene Chloride	4.32	84	20448m	0.344	ppbv	
26) Hexane	5.66	57	14988	0.107	ppbv #	45

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL040221\
Data File : VL036660.D
Acq On : 3 Apr 2021 00:51
Operator : SY/AP
Sample : M1853-03DL 40X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SOFF-SG-03DL

Manual Integrations
APPROVED

apatel
4/5/2021 1:14:17 PM

Quant Time: Apr 03 07:22:45 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar
QLast Update : Thu Mar 25 09:42:40 2021
Response via : Initial Calibration

