

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL040221\
 Data File : VL036663.D
 Acq On : 3 Apr 2021 2:44
 Operator : SY/AP
 Sample : M1834-02 1200X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV2

Manual Integrations
 APPROVED

apatel
 4/5/2021 1:14:22 PM

Quant Time: Apr 03 07:32:48 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar 25 2021
 QLast Update : Thu Mar 25 09:42:40 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.64	49	1334179	10.00	ppbv	-0.03
33) 1,4-Difluorobenzene	7.15	114	3372451	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.06	117	2996088	10.00	ppbv	-0.03

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.39	95	2390882	10.01	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.01	41	9291m	0.117	ppbv	
15) Acetone	3.85	43	128459	0.978	ppbv	97
20) Methylene Chloride	4.32	84	11620m	0.189	ppbv	
23) Vinyl Acetate	5.09	43	42521m	0.189	ppbv	
26) Hexane	5.67	57	15632m	0.108	ppbv	
34) 2-Butanone	5.24	43	81179m	0.493	ppbv	
44) 2,2,4-Trimethylpentane	7.84	57	87430	0.160	ppbv #	86
52) Tetrachloroethene	11.19	164	19838m	0.173	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL040221\
Data File : VL036663.D
Acq On : 3 Apr 2021 2:44
Operator : SY/AP
Sample : M1834-02 1200X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV2

Manual Integrations
APPROVED

apatel
4/5/2021 1:14:22 PM

Quant Time: Apr 03 07:32:48 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar
QLast Update : Thu Mar 25 09:42:40 2021
Response via : Initial Calibration

