

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040423\
 Data File : VL040305.D
 Acq On : 4 Apr 2023 14:52
 Operator : SY/MD
 Sample : QC033023 CAN 10158
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC033023 CAN 10158

Quant Time: Apr 05 01:30:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL031523AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Mar 15 16:14:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.182	49	997673	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.645	114	2520092	10.000	ppbv	-0.01
55) Chlorobenzene-d5	11.465	117	2222267	10.000	ppbv	-0.01

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.777	95	1844477	9.911	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL040423\
Data File : VL040305.D
Acq On : 4 Apr 2023 14:52
Operator : SY/MD
Sample : QC033023 CAN 10158
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC033023 CAN 10158

Quant Time: Apr 05 01:30:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL031523AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Mar 15 16:14:43 2023
Response via : Initial Calibration

