

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL040521\
 Data File : VL036687.D
 Acq On : 6 Apr 2021 00:27
 Operator : SY/AP
 Sample : M1945-02DL 40X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV2DL

Manual Integrations
 APPROVED

MMDadoda
 4/6/2021 5:14:34 PM

Quant Time: Apr 06 14:43:38 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar 25 2021
 QLast Update : Thu Mar 25 09:42:40 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1310811	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	3225290	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	2951816	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	2335258	9.93	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.87	43	114293m	0.886	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL040521\
Data File : VL036687.D
Acq On : 6 Apr 2021 00:27
Operator : SY/AP
Sample : M1945-02DL 40X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV2DL

Manual Integrations
APPROVED

MMDadoda
4/6/2021 5:14:34 PM

Quant Time: Apr 06 14:43:38 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar
QLast Update : Thu Mar 25 09:42:40 2021
Response via : Initial Calibration

