

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL040521\
 Data File : VL036697.D
 Acq On : 6 Apr 2021 10:29
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Quant Time: Apr 06 11:49:50 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar 25 09:42:40 2021
 QLast Update : Thu Mar 25 09:42:40 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1270235	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	3186871	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	2779413	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	2218565	10.02	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.20%

Target Compounds				Ovalue
15) Acetone	3.87	43	59531m	0.476 ppbv
20) Methylene Chloride	4.34	84	13331m	0.228 ppbv

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL040521\
Data File : VL036697.D
Acq On : 6 Apr 2021 10:29
Operator : SY/AP
Sample : VIBLK
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Apr 06 11:49:50 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar 25 09:42:40 2021
QLast Update : Thu Mar 25 09:42:40 2021
Response via : Initial Calibration

