

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL040521\
 Data File : VL036698.D
 Acq On : 6 Apr 2021 11:09
 Operator : SY/AP
 Sample : M1834-05DL2 120X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV5DL2

Manual Integrations
 APPROVED

MMDadoda
 4/6/2021 5:14:52 PM

Quant Time: Apr 06 11:52:45 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar 25 2021
 QLast Update : Thu Mar 25 09:42:40 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1236755	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	3134794	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	2845009	10.00	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.42	95	2231578	9.84	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.02	41	45768	0.621	ppbv	89
10) Heptane	8.12	43	208942	1.113	ppbv #	28
26) Hexane	5.69	57	294022	2.189	ppbv #	45
30) Cyclohexane	7.13	84	88717m	0.815	ppbv	
34) 2-Butanone	5.25	43	874686	5.716	ppbv	96
36) Benzene	6.89	78	168378	0.582	ppbv #	89
44) 2,2,4-Trimethylpentane	7.87	57	5521034	10.850	ppbv	97
53) Toluene	9.81	91	136638	0.444	ppbv	97
58) Ethyl Benzene	12.71	91	130143	0.314	ppbv	98
59) m/p-Xylene	12.97	91	158392m	0.439	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL040521\
Data File : VL036698.D
Acq On : 6 Apr 2021 11:09
Operator : SY/AP
Sample : M1834-05DL2 120X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV5DL2

Manual Integrations
APPROVED

MMDadoda
4/6/2021 5:14:52 PM

Quant Time: Apr 06 11:52:45 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar
QLast Update : Thu Mar 25 09:42:40 2021
Response via : Initial Calibration

