

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL040721\
 Data File : VL036742.D
 Acq On : 8 Apr 2021 1:18
 Operator : SY/AP
 Sample : M1929-09DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 FS-IA-03DL

Manual Integrations
 APPROVED

MMDadoda
 4/12/2021 11:16:38 AM

Quant Time: Apr 08 04:48:40 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Apr 08 04:48:40 2021
 QLast Update : Thu Apr 08 01:58:43 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1197646	10.00	ppbv	-0.01
33) 1,4-Difluorobenzene	7.17	114	2997155	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.08	117	2452772	10.00	ppbv	-0.01

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
68) 1-Bromo-4-Fluorobenzene	14.41	95	1919808	9.82	ppbv	-0.01	
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.20%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.02	41	24387m	0.342	ppbv	
13) Ethanol	3.62	45	25938m	3.053	ppbv	
15) Acetone	3.85	43	147375	1.250	ppbv #	78
19) Isopropyl Alcohol	3.99	45	76030m	1.173	ppbv	
20) Methylene Chloride	4.34	84	63712m	1.154	ppbv	
26) Hexane	5.68	57	38695	0.298	ppbv #	43
52) Tetrachloroethene	11.22	164	4347m	0.043	ppbv	
53) Toluene	9.80	91	214537	0.729	ppbv	97
59) m/p-Xylene	12.96	91	31593m	0.102	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL040721\
Data File : VL036742.D
Acq On : 8 Apr 2021 1:18
Operator : SY/AP
Sample : M1929-09DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
FS-IA-03DL

Manual Integrations
APPROVED

MMDadoda
4/12/2021 11:16:38 AM

Quant Time: Apr 08 04:48:40 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Apr
QLast Update : Thu Apr 08 01:58:43 2021
Response via : Initial Calibration

