

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL040821\
 Data File : VL036758.D
 Acq On : 8 Apr 2021 19:55
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

MMDadoda
 4/12/2021 11:47:18 AM

Quant Time: Apr 09 09:33:39 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar 25 2021
 QLast Update : Thu Mar 25 09:42:40 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1371414	10.00	ppbv	-0.01
33) 1,4-Difluorobenzene	7.17	114	3352516	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.08	117	2628063	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.41	95	1530610	7.31	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	73.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.02	41	7802m	0.096	ppbv	
15) Acetone	3.87	43	26461m	0.196	ppbv	
20) Methylene Chloride	4.33	84	14729m	0.233	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL040821\
Data File : VL036758.D
Acq On : 8 Apr 2021 19:55
Operator : SY/AP
Sample : VIBLK
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

MMDadoda
4/12/2021 11:47:18 AM

Quant Time: Apr 09 09:33:39 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL032521AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Mar
QLast Update : Thu Mar 25 09:42:40 2021
Response via : Initial Calibration

